

DECLARATION OF CONFORMITY TO COUNCIL DIRECTIVE 93/42/EEC CONCERNING MEDICAL DEVICES



MANUFACTURER:

CONTEC MEDICAL SYSTEMS CO., LTD

No.112 Qinhuang West Street, Economic & Technical
Development Zone, Qinhuangdao, Hebei Province,
PEOPLE'S REPUBLIC OF CHINA

MEDICAL DEVICE:

Electrocardiograph, ECG300G

CLASSIFICATION - ANNEX IX:

Class II a, Rule 10

CONFORMITY ASSESSMENT ROUTE: Annex II excluding chapter 4

WE, (CONTEC MEDICAL SYSTEMS CO., LTD) HERewith DECLARE THAT THE STATED
MEDICAL DEVICES. MEET THE TRANSPOSITION INTO NATIONAL LAW, THE PROVISIONS OF
COUNCIL DIRECTIVE 93/42/EEC OF 14 JUNE 1993 CONCERNING MEDICAL DEVICES;
INCLUDING, AT 21 MARCH 2010, THE AMENDMENTS BY COUNCIL DIRECTIVE 2007/47/EEC
ALL SUPPORTING DOCUMENTATION IS RETAINED AT THE PREMISES OF THE MANUFACTURE.

STANDARDS APPLIED: SEE ATTACHED LIST OF (HARMONISED - EN) STANDARDS FOR WHICH
DOCUMENTED EVIDENCE OF COMPLIANCE CAN BE PROVIDED.

NOTIFIED BODY:

TÜV SÜD PRODUCT SERVICE GMBH
RIDLERSTR 65, D-80339 MÜNCHEN, GERMANY

IDENTIFICATION NUMBER:

CE 0123

(EC) CERTIFICATE(S):

G1 050972 0050 Rev.02

EC REP

EUROPEAN REPRESENTATIVE:

Shanghai International Holding Corp. GmbH(Europe)
Eiffestrasse 80, 20537 Hamburg Germany

START OF CE-MARKING:

2010-03-20 (Date or Lot or serial number)

PLACE, DATE OF DECLARATION:

QINHUANGDAO, 2019-07-23

SIGNATURE:

 President

TF-CE090602.2-09

Ver: J

Page 1 of 2

DECLARATION OF CONFORMITY TO COUNCIL DIRECTIVE 93/42/EEC CONCERNING MEDICAL DEVICES

Appendix: list of (harmonised - EN) standards

| No. | Reference | Title |
|-----|-----------------------------------|--|
| 1 | EN 60601-1: 1990+A1:1993 +A2:1995 | Medical electrical equipment - Part 1: General requirements for safety |
| 2 | EN 60601-1-2: 2007 | Medical electrical equipment - Part 1-2: General requirements for basic safety and essential performance - Collateral standard: Electromagnetic compatibility - Requirements and tests |
| 3 | EN 60601-1-4:1996+A1: 1999 | Medical Devices Part 1-4: General Requirements for Safety - Programmable Medical Electrical Equipment |
| 4 | EN 60601-1-6:2007 | Medical electrical equipment – Part 1-6: General requirements for basic safety and essential performance - Collateral standard: usability |
| 5 | EN 60601-2-25:1995/A1:1999 | Medical electrical equipment - Part 2-25: Particular requirements for the safety of electrocardiographs |
| 6 | EN 62304: 2006 | Medical device software - Software life-cycle processes |